

## TAMILNADU WATER SUPPLY AND DRAINAGE BOARD

### DISTRICT WATER TESTING LABORATORY

No. 471 & 472, Bharathiyar Complex, Avinashi Road, Tiruppur – 641602.

Phone: 0421 - 2245201 E-mail: labtpr@gmail.com



TEST REPORT

Doc.No: DWTL/FORM/7.8/01

Sample / Lab No: 2711/2712 Report Issue Date: 12.11.2022
Invoice Details: 2779/Dt.11.11.2022/ Amt.Rs.1,652/-

Client Name & Address : M/s. Little Kingdom School, Contact No / Email.ID 74/1D, Solipalayam,

15 – Velampalayam,

Tiruppur.

Sample Name : Water

Sample Description : Bore well – RO water at M/s. Little Kingdom School

Sample Submitted by : Customer – M/s. Little Kingdom School

Location & Date : Above Address & 11.11.2022

Quantity of sample Received : 2 Litre

Sample Container : Plastic – Good condition

Sample Condition on Receipt : Good

Sample Received On : 11.11.2022
Test Commenced on : 11.11.2022
Test Completed on : 12.11.2022

Environmental Condition : 24.5°C

	,			Result IS 1	Specification as per IS 10500 : 2012	
S.No	Parameters	Test Method	Unit		Acceptable Limit	Permissible Limit in the absence of alternate source
1	pH @ 25 <sup>0</sup> С	1S 3025 Part 11-1983		6.95	6.5-8.5	6.5-8.5
2	Total Alkalinity as CaCO <sub>3</sub>	IS 3025–Part 23-1986	mg/L	19	200	600
3	Total Hardness as CaCO3	1S 3025 Part 21-2009	mg/L	18	200	600
4	Chloride as Cl	IS 3025-Part 32-1988	mg/L	8	250	1000
5	Sulphate as SO4	IS 3025-Part 24-1986	mg/L	5	200	400
6	Nitrate as NO <sub>3</sub>	APHA 23rd Edition 2017– 4500 –NO3 B	mg/L	Less than 2	45	45
7	Fluoride as F	APHA 23rd Edition 2017 – 4500- F-D	mg/L	Less than 0.2	1.0	1.5

B .George - Junior Water Analyst (A/c)

#### Note:

1. The test results relate only to the items tested.

2. The test report shall not be reproduced anywhere except in full and in the same format without the permission of the laboratory.

3. Unless informed by customer, the test items will not be retained for more than 15 days from the date of issue of the test report.


4. The result applied to the sample as received basis.

Issue No./Date	2711-2712/12.11.22	Amend, No/ Date	00/	Page No	1 of 2		



# TAMILNADU WATER SUPPLY AND DRAINAGE BOARD

## DISTRICT WATER TESTING LABORATORY





Phone: 0421 - 2245201

E-mail: labtpr@gmail.com

	IEST REF	ORI	Doc.No: DWTL/FORM/7.8/01		
Sample / Lab No:	2711/2712	Report Issue Date:	12.11.2022		
Invoice Details:	2779/Dt.11.11.2022/ Amt.Rs.1,652/-				

	Parameters	Test Method		Result	Specification as per IS 10500 : 2012	
S.No			Unit		Acceptable Limit	Permissible Limit in the absence of alternate source
8	Appearance	Physical observation		Clear	-	-1
9	Colour	IS 3025-Part 4	Hazen	Colourless	Agreeable	Agreeable
10	Odour	IS 3025-Part 5		None	-	-
11	Turbidity	IS 3025-Part 10	NTU	0	1	5
12	Total Dissolved Solids	IS 3025-Part 16	mg/L	52	500	2000
13	Conductivity	IS 3025-Part 14	μS/cm	74	-	-
14	P – Alkalinity as CaCO <sub>3</sub>	IS 3025–Part 23-1983	mg/L	0	-	-
15	Calcium as Ca	IS 3025-Part 40	mg/L	5	75	200
16	Magnesium as Mn	IS 3025–Part 46	mg/L	1	30	100
17	Sodium as Na	IS 3025–Part 45	mg/L	-:	-	-
18	Potassium as K	IS 3025–Part 45	mg/L	-		-
19	Iron as Fe	IS 3025–Part 53	mg/L	0.00	0.3	1
20	Manganese as Mn	IS 3025–Part 59	mg/L	0.00	0.1	0.3
21	Ammonia as NH <sub>3</sub>	IS 3025–Part 34	mg/L	0.00	0.5	0.5
22	Nitrite as NO <sub>2</sub>	IS 3025–Part 34	mg/L	0.00	-	-
23	Total Phosphate as PO <sub>4</sub>	IS 3025–Part 31	mg/L	0.00	-	-
24	Tidys 4 hr. test		mg/L	-	-	_
25	Fecal Coliform	IS 1622	CFU/ 100ml	0	0 / 100mL	0/100mL

....End of the report.....

Note: 1. The test results relate only to the items tested.

The test report shall not be reproduced anywhere except in full and in the same format without the permission of the laboratory.

3. Unless informed by customer, the test items will not be retained for more than 15 days from the date of issue of the test report. 4. The result applied to the sample as received basis.

Issue No./Date	2711-2712/12.11.22	Amend. No/ Date	00/	Page No	2 of 2

Authorised Signatory

B.George - Junior Water Analyst (A/c)